

## Certificate of Calibration

Reference No R123106148  
Date of Issue 16 Jun 2023  
Customer PRO LAB ENGINEERING SERVICES SDN BHD  
ID: 039228 No.21-g  
Jalan Bidara 8 saujana utama 3  
47000 Sungai Buloh  
Selangor  
Instrument Liquid In Glass Thermometer  
Model ZEAL  
Serial No 210403  
Control No CA8869J  
Equipment ID N/A  
Capacity/Range -10 °C to 250 °C  
Date of Receipt 15 Jun 2023  
Date of Calibration 16 Jun 2023  
Recalibration Date Customer to Determine  
(Specified by Customer) The User should be aware there are many factors may cause this instrument to drift out of calibration limits prior to the stated recalibration date.  
Condition of Instrument Before Calibration Good Physical Condition  
After Calibration Calibrated and Serviceable  
Location of Calibration Trescal Laboratory  
Calibration Environment (23 ± 2) °C, (55 ± 15) %rh  
Calibration Method LCP 01101

Cert. No. PSYP- 23042974

Page 1 of 2



### Reference Standard Used

Reference Instrument	Equipment ID	Control No	Certificate No	Traceable to	Due Date
Temperature Liquid Bath	PH-T-LB2	CI6300	PSYP-23010704	NMIM,NMIA	09 Feb 2024
Temperature Indicator With PRT Sensor	PH-T-PT14	CI6300S	PSYP-23004859	NMIM,NMIA	18 Jan 2024

Calibrated By



Muhammad Azlan Bin Zakaria

Approved Signatory



Jag Oh Joo Kiat

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

## Certificate of Calibration

Control No. CA8869J

Cert. No. PSYP-23042974

Page 2 of 2

### Instrument Calibrated

Resolution	1.0	°C
Readability	0.5	
Type	Mercury	
Immersion	76 mm	

Specification
± 1.0 °C

EMERGENT STEM TEMPERATURE	ACCURACY TEST			
	REFERENCE	UUT	CORRECTION	SPECIFICATION
	READ	READ		±
23	0.0	0.0	0.0	1.0
24	50.0	50.5	-0.5	
25	100.0	101.0	-1.0	
26	150.0	151.0	-1.0	
28	200.0	201.0	-1.0	
Measurement Uncertainty ± 0.5 °C			k = 2	

Info 1 : True Read = UUT Read + Correction

Info 2 : UUT - Unit Under Test

Info 3 : If the correction is out of user specification, in order to meet the specification the user shall apply correction to derive true value.

Info 4 : Calibration Curve can be derive by interpolation the calibration point, the interpolation point is valid through the linearity of curve.

Info 5 : Uncertainty ~ Parameter, associated with the result of measurement, that characterises the dispersion of the value that reasonably be attributed to the measurand.

Note 6 : Emergent Stem Temperature (EST) = Air temperature surrounding the part of the mercury column which are not immersed during calibration.

Note 7 : For more precision measurement, EST Correction shall added to 'Correction' where EST in use is not being same as calibrated EST.

Note 8 : Emergent Stem Temperature Correction =  $K N (t_c - t_u)$

where K : Coefficient of expansion of liquid; mercury  $K=0.00016 \text{ } ^\circ\text{C}^{-1}$ ; spirit  $K = 0.00096 \text{ } ^\circ\text{C}^{-1}$ ;

N : is the number of degrees exposed in air;

$t_c$  : Emergent stem temperature in calibration;

$t_u$  : Emergent stem temperature in use.

ACCEPTABLE TO USE

Izatul Azira  
Lab Manager  
Pro Lab Engineering Services Sdn Bhd

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

Lot 148, No. 2A, Jalan U1/19, Hicom-Glenmarie Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan, Malaysia. Tel: +603-5569 1648 Fax: +603-5569 1548

www.trescal.com

LQP 00115\_APA / 01 June 2022