

Certificate of Calibration

Reference No R122075556
Date of Issue 17 Jun 2022
Customer PRO LAB ENGINEERING SERVICES SDN BHD
ID: 039228 No.21-g
Jalan Bidara 8 saujana utama 3
47000 Sungai Buloh
Selangor
Instrument Hydrometer
Model G H ZEAL LTD
Serial No 18/646164
Control No CA0509F
Equipment ID N/A
Capacity/Range 0.995 g/mL to 1.038 g/mL
Date of Calibration 17 Jun 2022
Recalibration Date Customer to Determine
(Specified by Customer) The User should be aware there are many factors may cause this instrument to drift out of calibration limits prior to the stated recalibration date.
Condition of Instrument
Before Calibration Good Physical Condition
After Calibration Calibrated and Serviceable
Location of Calibration Trescal Laboratory
Calibration Environment (23 ± 2) °C, (55 ± 10) %rh
Calibration Method LCP 02101

Cert. No. PSYP- 22039112

Page 1 of 2



Reference Standard Used

Reference Instrument	Equipment ID	Control No	Certificate No	Traceable to	Due Date
Density Hydrometer	PH-DY-HY33	C9800L	NMIM-5124-T-21	NMIM	07 Dec 2023
Density Hydrometer	PH-DY-HY7	CU7598	NMIM-5123-T-21	NMIM	30 Nov 2022
Thermometer With Sensor	PH-V-TS2	CW2222	PSYP-21087201	NMIM	29 Dec 2022

Calibrated By



Nor Azfa Azlinda Binti Salleh

Approved Signatory



Fatimah Binti Azlan

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

Lot 148, No. 2A, Jalan U1/19, Hicom-Glenmarie Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan, Malaysia. Tel: +603-5569 1648 Fax: +603-5569 1548
www.trescal.com

Certificate of Calibration

Control No. CA0509F

Cert. No. PSYP-22039112

Page 2 of 2

Instrument Calibrated

Range	0.995 to 1.038	g/mL
Resolution	0.001	

ACCURACY TEST			
REFERENCE READ	UUT READ	CORRECTION	SPECIFICATION (\pm)
0.9950	0.9958	-0.0008	N/A
1.0165	1.0173	-0.0008	
1.0372	1.0380	-0.0008	
Measurement Uncertainty =		0.0006 g/mL	k = 2

Remark : (*) mean out of specification

Info :

1. Actual Reading ~ UUT read + Correction
2. Uncertainty ~ Parameter, associated with the result of a measurement, that characterizes the dispersion of the values that could reasonably be attributed to the measurand.
3. The calibration was done at the temperature of (23 ± 2 °C) and corrected to the standard temperature of hydrometer.

ACCEPTABLE TO USE



Izatul Azira
Lab Manager

Pro Lab Engineering Services Sdn Bhd

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

Lot 148, No. 2A, Jalan U1/19, Hicom-Glenmarie Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan, Malaysia. Tel: +603-5569 1648 Fax: +603-5569 1548
www.trescal.com