

Certificate of Calibration

Reference No R123107713

Cert. No. PSYP- 23048519

Date of Issue 11 Jul 2023

Page 1 of 2

Customer PRO LAB ENGINEERING SERVICES SDN BHD
No.21-g
Jalan Bidara 8 saujana utama 3
ID: 039228 47000 Sungai Buloh
Selangor



Instrument Water Bath
Model NL 2008/ 002
Serial No 0222040072
Control No CA6301J
Equipment ID N/A
Capacity/Range Max. 100 °C

Date of Receipt 05 Jul 2023

Date of Calibration 10 Jul 2023

Recalibration Date Customer to Determine

(Specified by Customer) The User should be aware there are many factors may cause this instrument to drift out of calibration limits prior to the stated recalibration date.

Condition of Instrument

Before Calibration Good Physical Condition

After Calibration Calibrated as Requested by Customer

Location of Calibration In-situ

Calibration Environment $(30 \pm 1) ^\circ\text{C}$, $(50 \pm 2) \% \text{rh}$

Calibration Method LCP 01102

Reference Standard Used

Reference Instrument

Temperature Recorder With Sensor

Equipment ID

PH-ST-RW18

Control No

C0002PS

Certificate No

PSYP-22051004

Traceable to

NMIM,NMIA

Due Date

18 Aug 2023

Calibrated By



Muhammad Iqbal Bin
Mohammad Fuad

Approved Signatory



Tamilselvam A/L Chinayah

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

Certificate of Calibration

Control No. CA6301J

Cert. No. PSYP-23048519

Page 2 of 2


Instrument Calibrated

Set	Resolution	0.1	°C
	Readability	0.1	
Calibration Area : 10% away from internal walls			

ACCURACY TEST

Set Temperature	37.8	60.0
Indicated Enclosure Temperature	37.8	60.0
Measured Enclosure Temperature	38.5	60.3
Spatial Variation	0.8	0.5
Temporal Variation	0.4	0.2
Overall Variation	1.0	0.7
Tolerance \pm	0.5	0.4

MAPPING			Measured Value		
Enclosure Position Code	1	Back	38.2	60.1	
	2	Back	38.1	60.1	
	3	Center	38.6	60.4	
* C * Center	4	Center	38.6	60.5	
	5	Front	38.9	60.5	
* A * Corner	6	Front	38.4	60.0	
* F * Center of Face	Measurement Uncertainty \pm 0.6 °C			k = 2	



Info 1 : Set Temperature ~ Temperature Controller Setting.

Info 2 : Indicated Enclosure Temperature ~ Temperature Indicator of Enclosure.

Info 3 : Measured Enclosure Temperature ~ Mid-range value of the temperature obtained from all relevant sites within the enclosure.

Info 4 : Spatial Variation ~ Difference between the mid-range of all measured temperature obtained at one site and at another site for those sites which give the greatest difference.

Info 5 : Temporal Variation ~ Maximum value of the temperature range obtained for each standard site throughout the test interval.

Info 6 : Overall Variation ~ The maximum variation of measure temperature in the working space.

ACCEPTABLE TO USE

Izatul Azira
Lab Manager

Pro Lab Engineering Services Sdn Bhd

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.