

Certificate of Calibration

Reference No R124135826
Date of Issue 12 Jun 2024
Customer PRO LAB ENGINEERING SERVICES SDN BHD
ID: 039228 No.21-G
Jalan Bidara 8 Saujana Utama 3
47000 Sungai Buloh
Selangor
Instrument Oven (Box Type)
Model NL1017X/009A
Serial No 0122040143
Control No CA4622J
Equipment ID OVEN 2
Capacity/Range Max. 300 °C
Date of Receipt 10 Jun 2024
Date of Calibration 11 Jun 2024
Recalibration Date 11 Jun 2025
(Specified by Customer) The User should be aware there are many factors may cause this instrument to drift out of calibration limits prior to the stated recalibration date.
Condition of Instrument
Before Calibration Good Physical Condition
After Calibration Calibrated and Serviceable
Location of Calibration In-situ
Calibration Environment (29.4 ± 0.6) °C, (42 ± 4) %rh
Calibration Method LCP 01102

Cert. No. PSYP- 24043760

Page 1 of 2



Reference Standard Used

Reference Instrument	Equipment ID	Control No	Certificate No	Traceable to	Due Date
Temperature Recorder With Sensor	PH-ST-RW4	CS9292S	PSYP-23042411	NMIM, NMIA	21 Jun 2024

Calibrated By



Muhammad Iqbal Bin
Mohammad Fuad

Approved Signatory



Tamilselvam A/L Chinayah

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the laboratory accreditation requirements of Skim Akreditasi Makmal Malaysia (SAMM) of Standards Malaysia which is a signatory to the ILAC MRA. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Metrology Institute of Malaysia (NMIM) and other recognised national metrology institutes. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

Certificate of Calibration

Control No. CA4622J

Cert. No. PSPP-24043760

Page 2 of 2


Controller	Resolution	0.1	°C
Setting	Readability	0.1	

Tolerance
N/A

PERFORMANCES (°C)		
Temperature Setpoint	105.0	110.0
Mean Indicated Temperature	105.0	110.0
Measured Highest Temperature	106.6	111.9
Measured Lowest Temperature	102.6	105.5
Measured Mean Temperature	104.9	110.0
Temperature Fluctuation	0.8	2.8
Temperature Gradient	3.3	5.5
Temperature Variation In Space	3.1	4.3

Measurement Point	Specified Location	Mean Temperature (°C)	
1	Top Corners	102.9	111.2
2		105.0	110.3
3		106.3	110.3
4		105.4	110.8
5	Centre of working space	106.0	110.3
6	Bottom Corners	104.8	111.5
7		106.0	111.1
8		103.8	108.1
9		103.7	106.0
Measurement Uncertainty, ± (k = 2)		1.0	1.4

Location of sensors



Size of chamber : 0.6 m x 0.5 m x 0.7 m

Volume of chamber : 210 liter

Sensor distance from wall (except no. 5) :
L/10 or minimum 50 mm, where L is the length,
width or height of the working space.

Info 1 : Temperature Setpoint : Desired temperature as set by the chamber controls.

Info 2 : Mean Indicated Temperature : Mean temperature reading of the chamber indicator.

Info 3 : Measured Highest Temperature : Highest temperature measured, after stabilization, from all the measurement points in the working space during a specified interval of time.

Info 4 : Measured Lowest Temperature : Lowest temperature measured, after stabilization, from all the measurement points in the working space during a specified interval of time.

Info 5 : Measured Mean Temperature : Mean value calculated from the mean temperature of all points.

Info 6 : Temperature Fluctuation : Greatest difference, after stabilization, between the maximum and minimum temperatures at specified points in the working space during a specified interval of time.

Info 7 : Temperature Gradient : Maximum difference in temperature value, after stabilization, at any moment in time between two separate points in the working space.

Info 8 : Temperature Variation In Space : Maximum difference in temperature value, after stabilization, at any moment in time between the temperature at the centre of the working space and at any other point in the working space.

ACCEPTABLE TO USE

Lab Manager/Approved Signatory
PRO LAB ENGINEERING SERVICES SDN BHD

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the laboratory accreditation requirements of Skim Akreditasi Makmal Malaysia (SAMM) of Standards Malaysia which is a signatory to the ILAC MRA. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Metrology Institute of Malaysia (NMIM) and other recognised national metrology institutes. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESICAL (MALAYSIA) SDN. BHD.

Lot 148, No. 2A, Jalan U1/19, Hicom-Glenmarie Industrial Park, 40150 Shah Alam, Selangor Darul Ehsan, Malaysia. Tel: +603-5569 1648 Fax: +603-5569 1548
www.trescal.com

LOP 00115_APA / 01 July 2023