

Certificate of Calibration

Reference No R123107865
 Date of Issue 11 Jul 2023
 Customer PRO LAB ENGINEERING SERVICES SDN BHD
 ID: 039228 No.21-g
 Jalan Bidara 8 saujana utama 3
 47000 Sungai Buloh
 Selangor
 Instrument Standard Weight
 Model N/A
 Serial No N/A
 Control No CA9989J
 Equipment ID N/A
 Capacity/Range 200 g
 Date of Calibration 11 Jul 2023
 Recalibration Date Customer to Determine
 (Specified by Customer) The User should be aware there are many factors may cause this instrument to drift out of calibration limits prior to the stated recalibration date.
 Condition of Instrument Before Calibration Good Physical Condition
 After Calibration Calibrated and Serviceable
 Location of Calibration Trescal Laboratory
 Calibration Environment (23 ± 2) °C, (55 ± 10) %rh
 Calibration Method LCP 01305

Cert. No. PSYP- 23048394

Page 1 of 2



Reference Standard Used

Reference Instrument	Equipment ID	Control No	Certificate No	Traceable to	Due Date
Weighing Comparator	PH-M-WC2	C6868	PSYP-22086694	NMIM	06 Dec 2023
Standard Weight	PH-M-SW1	CI0077	PSYP-23038821	NMIM	07 Jun 2024

Calibrated By

Nurzulaiha Binti Ahmad

Approved Signatory

Fatimah Binti Azlan

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.

Certificate of Calibration

Control No. CA9989J

Cert. No. **PSYP-23048394**

Page 2 of 2

Instrument Calibrated

Class :	F2	Material :	Stainless Steel	Pieces :	1
---------	----	------------	-----------------	----------	---

ACCURACY TEST					
REMARK	NOMINAL VALUE	CONVENTIONAL VALUE		UNCERTAINTY (±) k = 2	MPE (F2) (±)
		AFTER ADJUST	BEFORE ADJUST		
	200 g	-	200.00073 g	0.00100 g	0.00300 g

Info :

1. Nominal Value ~ A value use to designate a characteristic of a device or to give a guide to its intended use.
2. Conventional Value ~ The conventional mass value represents the mass value a weight of density 8000 kg/m³ which it air of density of 1.2 kg/m³, would balance the weight being calibrated.
3. Uncertainty ~ Estimated amount by which the observed or calculated value of a quantity may depart from the true value.
4. Before Adjust ~ Values found during initial calibration. '-' mean no adjustment.
5. Class ~ A class of weight which meet certain metrological requirements intended to keep the errors within specified limits.
6. MPE ~ Maximum Permissible Error. MPE with reference to OIML R111-1:2004(E).

ACCEPTABLE TO USE

Izatul Azira
Lab Manager

Pro Lab Engineering Services Sdn Bhd

The reported expanded measurement uncertainty is stated as the standard measurement uncertainty multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the SAMM which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realised to the corresponding national standards laboratory. Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the Head of the issuing laboratory.

TRESCAL (MALAYSIA) SDN. BHD.